

Correction

Correction: Dobbels, F., *et al.* How Accurate Are Electronic Monitoring Devices? A Laboratory Study Testing Two Devices to Measure Medication Adherence. *Sensors* 2010, 10, 1652-1660

Leentje De Bleser ¹, Sabina De Geest ^{1,2}, Sofie Vandenbroeck ¹, Johan Vanhaecke ³
and Fabienne Dobbels ^{1,*}

¹ Centre for Health Services and Nursing Research, Katholieke Universiteit Leuven, Kapucijnenvoer 35 box 7001, B-3000 Leuven, Belgium; E-Mails: Leentje.DeBleser@med.kuleuven.be (L.D.B.); Sofie.Vandenbroeck@med.kuleuven.be (S.V.)

² Institute of Nursing Science, University of Basel, Basel, Switzerland;
E-Mail: Sabina.DeGeest@unibas.ch

³ Heart Transplantation Program, University Hospitals of Leuven, Leuven, Belgium;
E-Mail: johan.vanhaecke@uz.kuleuven.be

* Author to whom correspondence should be addressed; E-Mail: fabienne.dobbels@med.kuleuven.be;
Tel: +32-16-336981; Fax: +32-16-336970.

Received: 18 May 2011 / Published: 19 May 2010

One of readers found typing errors made in Table 1, therefore, the table is corrected as follows,

Table 1. Results of the accuracy of the Helping Hand™ and the MEMS.

	HH without reminder	HH with reminder	MEMS 18 month battery	MEMS 2 year battery
Event level				
Missing registrations	5/420 (1.20%)	1/630 (0.16%)	33/1050 (3.14%)	0/1050 (0%)
Overregistrations	2/420 (0.48%)	1/630 (0.16%)	0/1050 (0%)	0/1050 (0%)
Device level				
Missing registrations	3/10 (30%)	1/15 (6.67%)	20/25 (80%)	0/25 (0%)
Overregistrations	2/10 (20%)	1/15 (6.67%)	0/25 (0%)	0/25 (0%)
100% correct functioning	7/10 (70%)	13/15 (87%)	5/25 (20%)	25/25 (100%)

References

1. De Bleser, L.; De Geest, S.; Vandenbroeck, S.; Vanhaecke, J.; Dobbels, F. How Accurate Are Electronic Monitoring Devices? A Laboratory Study Testing Two Devices to Measure Medication Adherence. *Sensors* **2010**, *10*, 1652-1660.

© 2011 by the authors; licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution license (<http://creativecommons.org/licenses/by/3.0/>).