

Supplementary Materials: High M-MDSC Percentage as a Negative Prognostic Factor in Chronic Lymphocytic Leukaemia

Michał Zarobkiewicz, Wioleta Kowalska, Sylwia Chocholska, Waldemar Tomczak, Agata Szymańska, Izabela Morawska, Agnieszka Wojciechowska and Agnieszka Bojarska-Junak

Table S1. The detailed configuration of cytometers.

BD FACS Canto II			
No	Laser	Filter setup	Detector name
1	405	450/50	PacificBlue
2		510/50	AmCyan
3		530/30	FITC
4	488	585/42	PE
5		670LP mirror	PerCp-Cy5.5
6		780/60	PE-Cy7
7	633	660/20	APC
8		780/60	APC-Cy7
BD FACS Aria IIu			
No	Laser	Filter setup	Detector name
1	488	530/30	FITC
2		576/26	PE
3		610/20	PE-TexasRed
4		695/40	PerCp-Cy5.5
5		780/60	PE-Cy7
6	633	660/20	APC
7		780/60	APC-Cy7



© 2020 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).