


Parameter	Units	Resolution	Accuracy	Range
pH	pH	0.01 pH	+/- 0.02 pH	0.00-14.00 pH
ORP	mV	0.1 mV	+/- 1.0 mV	+/- 2000.0 mV
Dissolved Oxygen	ppm	0.01 ppm	0-30 ppm: +/- 1.5%; 30-50 ppm: +/- 3%	0.00-50.00 ppm
Conductivity	mS/cm	0.1 mS/cm	+/- 1%	0.0-200.0 mS/cm
Total Dissolved Solids	ppm	1 ppm	+/- 1%	0-400,000 ppm
Salinity	PSU	0.01 PSU	+/- 2%	0.00 – 70.00 PSU
Turbidity	FNU	0.1 FNU	+/- 2%	0.0-1000.0 FNU



Supplementary Figure S1. Hanna 9289 Multiparameter Meter resolution and accuracy and probe position during measurement.