

## Article

# Towards an Optical Gas Standard for Traceable Calibration-Free and Direct NO<sub>2</sub> Concentration Measurements

Javis A. Nwaboh <sup>1,\*</sup>, Zhechao Qu <sup>1</sup>, Olav Werhahn <sup>1</sup> and Volker Ebert <sup>1,2,\*</sup>

<sup>1</sup> Physikalisch-Technische Bundesanstalt, Bundesallee 100, D-38116 Braunschweig, Germany; zhechao.qu@ptb.de (Z.Q.); Olav.Werhahn@ptb.de (O.W.)

<sup>2</sup> Technical University of Darmstadt, Department of Mechanical Engineering, Reactive Flows and Diagnostics, Otto-Berndt-Str. 3, 64287 Darmstadt, Germany

\* Correspondence: javis.nwaboh@ptb.de (J.A.N.); volker.ebert@ptb.de (V.E.)

**Abstract:** We report a direct tunable diode laser absorption spectroscopy (dTDLAS) instrument developed for NO<sub>2</sub> concentration measurements without chemical pre-conversion, operated as an Optical Gas Standard (OGS). An OGS is a dTDLAS instrument that can deliver gas species amount fractions (concentrations), without any previous or routine calibration, which are directly traceable to the international system of units (SI). Here, we report NO<sub>2</sub> amount fraction quantification in the range of 100–1000 μmol/mol to demonstrate the current capability of the instrument as an OGS for car exhaust gas application. Nitrogen dioxide amount fraction results delivered by the instrument are in good agreement with certified values of reference gas mixtures, validating the capability of the dTDLAS-OGS for calibration-free NO<sub>2</sub> measurements. As opposed to the standard reference method (SRM) based on chemiluminescence detection (CLD) where NO<sub>2</sub> is indirectly measured after conversion to NO, titration with O<sub>3</sub> and the detection of the resulting fluorescence, a dTDLAS-OGS instrument has the benefit of directly measuring NO<sub>2</sub> without distorting or delaying conversion processes. Therefore, it complements the SRM and can perform fast and traceable measurements, and side-by-side calibrations of other NO<sub>2</sub> gas analyzers operating in the field. The relative standard uncertainty of the NO<sub>2</sub> results reported in this paper is 5.1% ( $k = 1$ , which is dominated (98%) by the NO<sub>2</sub> line strength), the repeatability of the results at 982.6 μmol/mol is 0.1%, the response time of the instrument is 0.5 s, and the detection limit is 825 nmol/mol at a time resolution of 86 s.

**Keywords:** gas analysis; laser absorption spectroscopy; dTDLAS; nitrogen dioxide (NO<sub>2</sub>); metrology



**Citation:** Nwaboh, J.A.; Qu, Z.; Werhahn, O.; Ebert, V. Towards an Optical Gas Standard for Traceable Calibration-Free and Direct NO<sub>2</sub> Concentration Measurements. *Appl. Sci.* **2021**, *11*, 5361. <https://doi.org/10.3390/app1125361>

Academic Editor: Vincenzo Spagnolo

Received: 30 April 2021

Accepted: 1 June 2021

Published: 9 June 2021

**Publisher's Note:** MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



**Copyright:** © 2021 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

## 1. Introduction

Nitrogen dioxide (NO<sub>2</sub>) is a prominent air pollutant emitted from combustion processes, burning fuel, cars or power plants. Exposure to elevated levels of NO<sub>2</sub> in the atmosphere causes serious respiratory diseases [1]. Accurate and reliable measurement methods for amount fraction (concentration) levels below 1 ppm are needed for atmospheric NO<sub>2</sub> air quality control [2] and to monitor NO<sub>2</sub> emitted directly from the exhausts of cars (1–2500 μmol/mol, note: μmol/mol is often called ppm) as required for the type approval emission limits quality control test (i.e., the EURO 5/6 emission standard states the range of NO<sub>2</sub> emission levels, which need to be measured during type approval, to be in the 1–30 μmol/mol for cars with a functional exhaust catalyst, and 30–1000 μmol/mol for cars with a malfunctioning exhaust catalyst). The standard reference method (SRM) for measuring NO<sub>2</sub> is based on chemiluminescence detection (CLD), as documented in the European standard EN 14211:2012 [3], where NO<sub>2</sub> is only measured *indirectly* (i.e., NO<sub>2</sub> in a sample is calculated from the difference of NO<sub>x</sub> and NO after the conversion (catalytic or photolytic) of NO<sub>2</sub> to NO. The NO then reacts with the ozone (O<sub>3</sub>) to produce electronically excited NO\*<sub>2</sub> which decays to NO<sub>2</sub> and emits light (luminescence) that is detected by, e.g., a photomultiplier tube. The NO<sub>x</sub> concentration is determined using the

measured signals, and the NO<sub>2</sub> concentration by subtracting NO (detected in an independent parallel channel) from NO<sub>x</sub> [4,5]). This indirect measurement of NO<sub>2</sub> can lead to an overestimation of NO<sub>2</sub>, as other NO<sub>y</sub> gas species (HNO<sub>3</sub>, HNO<sub>2</sub>, N<sub>2</sub>O<sub>5</sub>) in the sample can also be converted into NO as well. It is worth noting that NO<sub>2</sub> is the only air pollutant where the SRM is based on an *indirect* measurement. Due to the fact that NO<sub>2</sub> is only indirectly measured, very reactive and that there is a lack of accurate static calibration reference materials CRM (gases), e.g., for car exhaust gas applications, in the 1–2500 ppm range with  $\leq 1\%$  uncertainties, there is the need to implement new accurate and reliable measurement methods (reference methods for ambient air and other applications) for *direct* NO<sub>2</sub> measurements [2] and alternative calibration approaches (e.g., an instrument standard OGS) to complement CRMs [6].

Laser spectroscopic techniques provide an option to develop opportunities for accurate and reliable measurement methods for direct NO<sub>2</sub> measurements [2]. Laser absorption spectroscopy has been demonstrated for direct measurements of the concentration of a variety of molecules in different applications such as atmospheric monitoring, combustion diagnostics and breath analysis [7–21], and therefore can serve as a good candidate to develop a reference method for direct NO<sub>2</sub> measurements. There have been several studies to perform direct NO<sub>2</sub> measurements and also addressing the shortcomings in the SRM-based measurements [2,22–26]. For atmospheric NO<sub>2</sub> measurements, spectroscopic techniques such as tunable diode laser absorption spectroscopy (TDLAS), quantum cascade laser absorption spectroscopy (QCLAS), cavity ringdown spectroscopy (CRDS), photoacoustic spectroscopy (PAS) and cavity attenuated phase shift spectroscopy (CAPS) have been reported with sub ppb detection limits [22,24,27–29]. Differential optical absorption spectroscopy (DOAS), TDLAS, PAS and CRDS have been successfully demonstrated for the detection of NO<sub>2</sub> emitted from automobile and aircraft engines, with detection limits in ppm down to the ppb range [23,30–33]. To date, for these spectroscopic NO<sub>2</sub> measurements, information on the metrological traceability and uncertainty assessment following, e.g., the guide to the expression of uncertainty in measurement (GUM) [34] is rarely reported. Metrological traceability of results (i.e., traceability of the results to the SI) and uncertainty assessments according to the GUM [34] of the NO<sub>2</sub> results derived using spectroscopic measurement methods are critical for quality control measurements [2] and type approval purposes [35].


Direct tunable diode laser absorption spectroscopy (dTDLAS) combined with the TILSAM (traceable infrared laser-spectrometric amount fraction measurement) method [36], has been demonstrated in a variety of applications addressing metrology data quality (SI-traceability and GUM compliant uncertainty assessments). dTDLAS is a variant of TDLAS that combines TDLAS with a special data evaluation procedure to derive concentration results that are directly traceable to the SI [7,37]. As reported on our previous work, a dTDLAS instrument is referred to as an optical gas standard (OGS) [7–10,37]. An OGS is similar to the ozone standard reference photometer (SRP) [38]. Employing dTDLAS, fast measurement frequencies (from sub-Hz to Hz) and response times (e.g., milliseconds to seconds) required in field measurements can be reached [39–42]. Focusing here on a new development for direct NO<sub>2</sub> (reactive gas) measurements in car exhaust gases, our previous works in [7–10,37] focused on dTDLAS instrument developments for the measurement of H<sub>2</sub>O, CO and CO<sub>2</sub> and H<sub>2</sub>O as well as reaching fast measurement frequencies [39–42].

In this paper, we report the details of a new dTDLAS spectrometer [43] specifically designed and developed for accurate direct NO<sub>2</sub> measurements. The focus of this work was to report the development stages of the NO<sub>2</sub> dTDLAS instrument towards an OGS (an instrument standard for direct NO<sub>2</sub> measurements), the testing and validation steps performed, the potentials and the current limitations of the instrument. The capabilities for direct NO<sub>2</sub> measurements (looking at all measured quantities desired for the results) are demonstrated via NO<sub>2</sub> concentration measurements. A comparison of calibration-free dTDLAS NO<sub>2</sub> concentration results (range: 100–1000  $\mu\text{mol/mol}$ , within the car exhaust NO<sub>2</sub> concentration range) versus certified reference gas mixtures is done as a validation

step. We further show that the instrument is capable of direct NO<sub>2</sub> measurements in car exhaust gas matrix. In order to compete with, and compliment CRMs and supplement the SRM for NO<sub>2</sub> measurements, with other gas matrices such as the car exhaust gas matrix in focus, the traceability of the spectroscopic results to the SI is addressed and an uncertainty assessment following GUM principles is made for the NO<sub>2</sub> results.


## 2. Experimental

Figure 1 depicts a schematic of the dTDLAS spectrometer for direct NO<sub>2</sub> measurements [43]. The instrument employs an interband cascade DFB laser (Nanoplus) emitting at 3.4 μm, a 77 cm silconert® coated stainless-steel single pass gas cell and a photovoltaic mid-infrared detector (Vigo system, detector diameter: 1 mm<sup>2</sup>). The sampling lines as well as the gas cell are coated to minimize the potential adsorption effects of NO<sub>2</sub> in the gas sample. For spectroscopic measurements, NO<sub>2</sub> gas samples flow (typical flow rate: 1.5 L/min) through the gas cell controlled by a mass flow controller (MFC, Bronkhorst) at the gas input and a valve at the output (connected to a pump) of the sampling lines as shown in Figure 1. The laser wavelength is swept at a frequency of 139.8 Hz (similar to [7–10]), and the collimated laser light is passed through the gas cell on to the detector. The signal from the detector, pre-amplified with a transimpedance amplifier (Femto DLPCA 200), is acquired using a data acquisition card (DAC, 18 bits, 625 KS/s NI PCI-6289) and is subsequently saved on a personal computer (PC) with a state-of-the-art operating system.



**Figure 1.** Schematic of the dTDLAS instrument for direct NO<sub>2</sub> measurements (Det: detector).

Figure 2a depicts typical signals (with NO<sub>2</sub> absorption lines visible around 2919.6 cm<sup>-1</sup>) measured at 0.2 Hz (the measurement frequency for single scan measurements is 18 Hz, i.e., ~0.056 s time resolution which is longer than the acquisition time of approximately 0.00352 s (2200 points on the up ramp at 625KS/s) due to the extra time needed to save the data using our acquisition software and corresponds to a time resolution of about ~5.6 s, i.e., ~0.2 Hz measurement frequency for 100 averaged single scans) for a gas mixture with an NO<sub>2</sub> amount fraction of 979 μmol/mol in the air. It could be noted here that the measurement frequency is different from the laser modulation frequency of 139 Hz. The absorbance data (Figure 2b derived using the data in Figure 2a, 100 averaged scans, 739 points in the fit window) are fitted by a set of 8 Voigt profiles [44] in order to derive the line areas (areas underneath the absorption lines), i.e., together with the NO<sub>2</sub> line at 2919.59 cm<sup>-1</sup>, 7 other neighboring NO<sub>2</sub> lines (see right y axis and the lines in red: line strength [45] plot in Figure 2b) are fitted in the spectral window selected for data analysis. The fitting was performed with the line positions, the line areas and the Lorentzian line widths kept free. The Doppler line widths were calculated at the measured gas temperature.



**Figure 2.** (a) Typical waves at 108 hPa signals measured using the spectrometer in Figure 1; (b) absorbance data derived from the data in Figure 2a. With the eight NO<sub>2</sub> lines shown (in red) in the line strength plot (right y axis), eight Voigt profiles are fitted to the data in order to derive the line areas.

The bottom panel of Figure 2b shows the residuals from the multiline Voigt fit with  $1\sigma_{\text{total}} = 2.7 \times 10^{-4}$ . Using the  $1\sigma_{\text{total}}$  value and the peak absorbance value of 0.092, a signal-to-noise ratio of 341 was calculated for the data in Figure 2b. Due to the signal-to-noise ratio (SNR) around and below 100 and the complex multi-line structure of the NO<sub>2</sub> line used, higher order line shape profiles like Galatry or Hartmann–Tran profiles [46–48] were not found to be suitable and advised to fit the data in this work. Using the line area of the NO<sub>2</sub> line at 2919.59 cm<sup>-1</sup>, derived for the data in Figure 2b, the NO<sub>2</sub> amount fraction (concentration)  $x_{\text{NO}_2}$  can be derived using Equation (1) (which is derived from the Beer Lambert law) [7]:

$$x_{\text{NO}_2} = \frac{k_B \cdot T}{S_T \cdot L \cdot p_{\text{total}}} \int_{-\infty}^{\infty} A(\tilde{\nu}) d\tilde{\nu} = \frac{k_B \cdot T}{S_T \cdot L \cdot p_{\text{total}}} \cdot A_{\text{line}} \quad (1)$$

where the quantity  $k_B$  is the Boltzmann constant,  $S_T$  is the line strength of the probed molecular transition at gas temperature  $T$ ,  $L$  is the optical path length of the light beam transmitted through the absorbing medium,  $p_{\text{total}}$  the total gas pressure,  $A(\tilde{\nu})$  the spectral absorbance and  $A_{\text{line}}$  the line area. The amount fraction (concentration) results  $x_{\text{NO}_2}$  is directly traceable to the SI if all the quantities in the right-hand side of Eqn. 1 are traceable to the SI as well [7–9,37]. To this end,  $x_{\text{NO}_2}$  results are derived without the calibration of the instruments with a reference gas mixture. As introduced, an instrument that can deliver amount fraction results that are directly traceable to the SI shall be referred to as an optical gas standard (OGS).


From the data in Figure 2b, a line area value of 0.0010999414 cm<sup>-1</sup> is derived for the NO<sub>2</sub> line 2919.59 cm<sup>-1</sup>, and a dTDLAS amount fraction value of (982.0 ± 49.1) μmol/mol calculated according to Eqn. 1 is combined with the TILSAM method [36]. This measured NO<sub>2</sub> amount fraction is in good agreement with the certified value of (979 ± 20) μmol/mol reported for the probed gas mixture, demonstrating the capability of the instrument for an absolute (“calibration-free”) direct NO<sub>2</sub> amount fraction measurements.

Table 1 holds an uncertainty budget for dTDLAS NO<sub>2</sub> results. As shown in Table 1, the combined uncertainty of the NO<sub>2</sub> results is 5.1% ( $k = 1$ ). The quantity  $k$  is a coverage factor [34]. The uncertainty of the amount fraction results of 5.1% is dominated by the uncertainty of the measured line area (1%) and the literature line strength value (5% [45]) as indicated by the higher index values (% individual contribution) in Table 1. The uncertainties of the other significant input quantities ( $p_{\text{total}}$ ,  $T$ ,  $k_B$ ,  $L$ ) are all less than 0.2%.

**Table 1.** Uncertainty budget of the NO<sub>2</sub> amount fraction results.

Parameter	Value	Relative Standard Uncertainty ( $k = 1$ )/%	Index (% Individual Contribution)
Pressure	108.5 hPa	0.20	0.10
Temperature	294.3 K	0.10	≤0.05
Path length	77.4 cm	0.13	≤0.05
Line strength (HITRAN value [40])	$5.425 \cdot 10^{-21}$ cm/molecule	5.00	96.00
Line area	$0.0010999414 \text{ cm}^{-1}$	1.00	3.80
NO <sub>2</sub> concentration ( $x_{\text{NO}_2}$ ) result	982 $\mu\text{mol/mol}$	5.10 (combined uncertainty)	


In order to determine the repeatability of the measurements delivered by the instrument, Figure 3a shows repeated measurements of NO<sub>2</sub> amount fractions performed with the 979  $\mu\text{mol/mol}$  NO<sub>2</sub> in synthetic air gas mixture. The mean value of the results in Figure 3a is 982.6  $\mu\text{mol/mol}$ , the reproducibility ( $1\sigma$ , standard deviation) is 8.7  $\mu\text{mol/mol}$  (0.9%) and the repeatability (standard deviation of the mean) is 0.5  $\mu\text{mol/mol}$  (0.1%). The excellent repeatability of 0.1% demonstrates the capability of a dTDLAS instrument for NO<sub>2</sub> OGS applications. The inset in Figure 3a shows a histogram of the measurements, depicting a normal distribution of the results with a full width at half maximum (FWHM) of 9.2  $\mu\text{mol/mol}$  (corresponding to an SNR of 107 for the amount fraction results at 979  $\mu\text{mol/mol}$ ). Figure 3b depicts a plot of the Allan deviation [49] of the NO<sub>2</sub> amount fraction results (here, the measurements are done at 2 Hz, i.e., at 0.5 s with 10 averaged scans) as a function of time. From the Allan plot in Figure 3b, an optimal precision of 825 nmol/mol (detection limit) is evaluated (conservatively, as there is no minimum visible) for the instrument at a time resolution of 86 s.



**Figure 3.** (a) Repeated NO<sub>2</sub> measurements: plot of NO<sub>2</sub> OGS amount fraction results vs. measurement number; (b) Allan deviation: plot of the  $x_{\text{NO}_2}$  Allan deviation as a function of time.

Repeated measurements were performed at different amount fraction levels (switching from about 100 to 1000  $\mu\text{mol/mol}$  as shown on Figure 4a) to study the behavior, response and linearity of the instrument. The gas mixture used for the measurements at the lowest amount fraction level of 100  $\mu\text{mol/mol}$  was a primary gas mixture (SI traceable, prepared using gravimetry by a national metrology institute). At 300  $\mu\text{mol/mol}$  we used a commercial gas mixture (secondary: SI-traceable, with a Deutscher Kalibrierdienst-DKD certificate [50]) and that at 1000  $\mu\text{mol/mol}$  was a commercial gas mixture from the same manufacturer (internally assigned, with no SI-traceability information provided). The

standard uncertainty reported for the NO<sub>2</sub> amount fractions in these gas mixtures are all 1% ( $k = 1$ ), corresponding to expanded uncertainties of 2% ( $k = 2$ ).



**Figure 4.** (a) Plot of the NO<sub>2</sub> OGS amount fraction results at different levels as a function of time (at  $\Delta t = 86$  s,  $1\sigma$  for 108.4, 305.0 and 983.5 are 6.2, 9.9 and 9.6, respectively); (b) dTDLAS NO<sub>2</sub> amount fraction  $x_{\text{NO}_2}$  results as a function-certified values of gas mixtures.


As presented, the sampling lines and the gas cell were all coated by Silconert, resulting in a fast response time of 0.5 s for the NO<sub>2</sub> measurements in Figure 4a. It could be noted here that this response time of 0.5 s does not also include the cylinder switching time and the time of manually opening valves to introduce the NO<sub>2</sub> gas sample into the gas cell. The repeatability (calculated as the standard deviation of the mean) at the lowest NO<sub>2</sub> results at the 100  $\mu\text{mol/mol}$  level is 0.5  $\mu\text{mol/mol}$  (0.5% relative). At the respective amount fraction levels,  $1\sigma$  (standard deviation) values were evaluated, demonstrating the short-term reproducibility ( $<10$   $\mu\text{mol/mol}$  at  $1\sigma$ ) for the NO<sub>2</sub> OGS results. Using the respective  $1\sigma$  values, an SNR from 18 at 100  $\mu\text{mol/mol}$  to 106 at 1000  $\mu\text{mol/mol}$  levels, at a time resolution of 86 s, were evaluated for the measurements in Figure 4a.

Figure 4b shows a plot of the NO<sub>2</sub> amount fraction results as a function of the certified values reported for the probed gas mixtures. A generalized linear regression (GLR) was applied to the data in Figure 4b, resulting in a slope of  $0.984 \pm 0.095$  ( $k = 2$ ) and an intercept value of  $(10.25 \pm 16.10)$   $\mu\text{mol/mol}$ , respectively. The slope value of 0.984 shows the good accuracy of the dTDLAS instrument, while the small deviation from the linear fit shows good linearity. The intercept value is consistent with zero, indicating an offset free measurement. The relative residuals from the linear model in the bottom panel of Figure 4b are all within the 1% range and thus well within the expanded uncertainty of the dTDLAS results of 5.1%,  $k = 1$ , which shows good agreement (compare well; validation step) with the certified values associated with the gas mixtures.

### 3. Discussion

The NO<sub>2</sub> OGS spectrometer was tested in the concentration range from 100 to 1000  $\mu\text{mol/mol}$  and was found to be in good agreement with the certified values of the reference gases, demonstrating the calibration-free OGS capability of the instrument. The detection limits of the dTDLAS instrument are 0.825  $\mu\text{mol/mol}$  (at  $\Delta t = 86$  s) and the relative uncertainty of the results delivered by the instrument is 5.1%,  $k = 1$ . This relative standard uncertainty was derived following GUM [34] principles. These results indicate that the instrument has the potential to measure NO<sub>2</sub> (that can be in the 1 to 1000  $\mu\text{mol/mol}$ : detection limit of 0.825  $\mu\text{mol/mol}$ ) emitted from the exhaust of cars as further supported by the simulation (HITRAN [45] line data used) in Figure 5 where the interference (due to absorption by other exhaust gas molecules) on NO<sub>2</sub> absorption is insignificant (absorbance

in the order of  $10^{-5}$  for H<sub>2</sub>O, NO<sub>2</sub>, CO<sub>2</sub>, and CO:  $L = 77.4$  cm,  $p = 100$  hPa,  $T = 296$  K). As presented, the relative repeatabilities of the NO<sub>2</sub> results at 1000  $\mu\text{mol/mol}$  and lowest amount fraction of 100  $\mu\text{mol/mol}$  are 0.1% and 0.5%, respectively. Table 2 holds a summary of the capabilities of the NO<sub>2</sub> OGS instrument presented in this work.



**Figure 5.** Simulation: NO<sub>2</sub> in exhaust gas matrix.

**Table 2.** Summary of NO<sub>2</sub> OGS instrument capabilities.

Quantity	Value
Current concentration range (NO <sub>2</sub> )	100–1000 $\mu\text{mol/mol}$
Optimal precision	0.825 $\mu\text{mol/mol}$ at $\Delta t = 86$ s
GUM compliant total relative uncertainty	5.1% ( $k = 1$ )
Relative repeatability at 1000 $\mu\text{mol/mol}$ and at 100 $\mu\text{mol}$ levels	0.1% and 0.5%
Time resolution (max = single scan to 10 scan average)	0.06 * sec to 0.5 s

\* >acquisition time of 3.5 ms as extra time is needed to save the data.

The relative uncertainties of the NO<sub>2</sub> results of 5.1% ( $k = 1$ ) are dominated (96% of the total uncertainty) by the uncertainty of the line strength value. Therefore, to further reduce the uncertainty of the results, future work will be focused on reducing the uncertainty of the line strength value. Since the dTDLAS results are in good agreement with the certified values (see Figure 4b), it can be concluded that the line strength used for the analysis is validated. The line strength value is taken from the HITRAN database [45]. This first validation of the line strength value was further assessed in a separate analysis to determine the line strength value from the data measured with the 300  $\mu\text{mol/mol}$  gas mixture, resulting in a line strength value of  $(5.444 \pm 0.076) \cdot 10^{-21}$  cm/molecule,  $k = 1$ , at 294.4 K for the NO<sub>2</sub> line 2919.59  $\text{cm}^{-1}$ . It could be noted here that the accuracy of the measured line strength would be dependent on the accuracy of the certified gas mixture used. This line strength value  $(5.444 \pm 0.076) \cdot 10^{-21}$  cm/molecule is in agreement with HITRAN's [45] value of  $(5.425 \pm 0.271) \cdot 10^{-21}$  cm/molecule,  $k = 1$ , at 296 K, further validating the HITRAN value [45]. In order to test the effect of a new line strength value with a reduced uncertainty, we used the line strength value  $5.444 \cdot 10^{-21}$  cm/molecule and the standard uncertainty  $(0.076 \cdot 10^{-21}$  cm/molecule: 1.4% relative) to evaluate the uncertainty of the NO<sub>2</sub> results (similar to Table 1), yielding relative uncertainty of 1.7% ( $k = 1$ ). The uncertainty of the dTDLAS results currently stands at 5.1%, and a 1.7% relative uncertainty for the NO<sub>2</sub> amount fraction results will fulfill the requirements of portable

emission measurement systems (PEMS) of less than 2% for NO<sub>2</sub>. Regarding the line strength, we estimate that we can achieve an uncertainty better than 1.7% or 0.2% in future line strength measurements, similarly to in our previous works [51,52], which would lead to a corresponding improvement in the NO<sub>2</sub> amount fraction uncertainty. One of the critical points in achieving an OGS is ensuring direct traceability to the SI. The traceability of the NO<sub>2</sub> amount fraction results in this work was addressed via the traceability of the input quantities on the right-hand side of Eqn.1, such as quantities  $p_{\text{total}}$ ,  $T$ ,  $L$  (SI-traceable to respective PTB standards). As presented, the validated line strength  $S_T$  value was taken from the HITRAN database [45] and SI-traceability was not reported. Therefore, in addition to a reduced uncertainty in the line strength value, new SI-traceable NO<sub>2</sub> line strength measurements are planned in separate experiments similar to [51,52].

This work shows that an OGS for NO<sub>2</sub> amount fraction measurements is feasible. An OGS (providing *direct* NO<sub>2</sub> amount fraction results traceable to the SI) for NO<sub>2</sub> can complement the current standard reference method (based on CLD), where NO<sub>2</sub> is “only” *indirectly* measured (note: the concept of indirect measurement is not the issue here, but rather the conversion efficiency of NO<sub>2</sub> to NO and maintaining a minimum residence time). By providing amount fraction results that are directly traceable to the SI, an OGS can also complement (use in the place of) calibrate reference gases both in the laboratory and in the field. An OGS will add the advantage of simultaneous in situ measurement and side-by-side field NO<sub>2</sub> calibrations of other analyzers.

#### 4. Conclusions

We presented a new dTDLAS NO<sub>2</sub> instruments are designed to be operated as an optical gas standard (OGS) and can thus perform absolute (“calibration free”) NO<sub>2</sub> amount fraction measurements directly, i.e., without any chemical conversion which is needed in the current standard reference method CLD. Nitrogen dioxide amount fractions in the range of 100–1000 µmol/mol were delivered by the instrument and are in good agreement with certified reference values attributed to the reference gas mixtures. The relative uncertainty of the results is currently at 5.1% ( $k = 1$ ) and is almost completely (96%) dominated by the inaccurate molecular reference data. The relative repeatability of the results at 982.6 µmol/mol is 0.1% and the instrument response time ranges from 0.5 s with a 10-scan average to 0.06 s in the single scan mode. Future work will focus on measuring a new SI-traceable NO<sub>2</sub> line strength value to further reduce the line strength uncertainty influence and thus the instrument accuracy. The instruments’ detection limit of 0.825 µmol/mol makes it well suited for direct measurements of the NO<sub>2</sub> concentration in car exhaust gases. Therefore, further plans are to extend the capability of the instrument for simultaneous measurements of NO<sub>2</sub> and NO concentrations (in both atmospheric, i.e., less than 1 µmol/mol level, and car exhaust gas species).

**Author Contributions:** Experiment concept: J.A.N., Z.Q., V.E. Experiment realization and data collection: J.A.N., Z.Q. Data evaluation and discussion: J.A.N., Z.Q., O.W., V.E. Paper conceptualization writing—review and editing: J.A.N., Z.Q., O.W. and V.E. All authors have read and agreed to the published version of the manuscript.

**Funding:** This work was internally funded.

**Institutional Review Board Statement:** Not applicable.

**Informed Consent Statement:** Not applicable.

**Acknowledgments:** The authors would like to acknowledge the support from the PTB working group 3.42 (Spectrometric Gas Analysis) in the development of the instrument.

**Conflicts of Interest:** The authors declare no conflict of interest.

## References

1. Salonen, H.; Salthammer, T.; Morawska, L. Human exposure to NO<sub>2</sub> in school and office indoor environments. *Environ. Int.* **2019**, *130*, 104887. [CrossRef] [PubMed]
2. EMPIR Project 16ENV02 Metrology for Nitrogen Dioxide (MetNO2). Available online: <http://empir.npl.co.uk/metno2/> (accessed on 29 April 2021).
3. European Standard: Ambient Air—Standard Method for the Measurement of the Concentration of Nitrogen Dioxide and Nitrogen Monoxide by Chemiluminescence, EN 14211:2012. Available online: <https://www.beuth.de/en/standard/din-en-14211/146720023> (accessed on 27 May 2021).
4. Miller, C.C. Chemiluminescence Analysis and Nitrogen-Dioxide Measurement. *Lancet* **1994**, *343*, 300–301. [CrossRef]
5. Steinbacher, M.; Zellweger, C.; Schwarzenbach, B.; Bugmann, S.; Buchmann, B.; Ordóñez, C.; Prevot, A.S.H.; Hueglin, C. Nitrogen oxide measurements at rural sites in Switzerland: Bias of conventional measurement techniques. *J. Geophys. Res. Atmos.* **2007**, *112*, D11307. [CrossRef]
6. EMPIR Project 19ENV09 Improved Vehicle Exhaust Quantification by Portable Emission Measurement Systems Metrology (MetroPEMS). Available online: <https://www.metropems.ptb.de/home/> (accessed on 29 April 2021).
7. Nwaboh, J.A.; Qu, Z.; Werhahn, O.; Ebert, V. Interband cascade laser-based optical transfer standard for atmospheric carbon monoxide measurements. *Appl. Opt.* **2017**, *56*, E84–E93. [CrossRef]
8. Buchholz, B.; Böse, N.; Ebert, V. Absolute validation of a diode laser hygrometer via intercomparison with the German national primary water vapor standard. *Appl. Phys. B* **2014**, *116*, 883–899. [CrossRef]
9. Pogany, A.; Wagner, S.; Werhahn, O.; Ebert, V. Development and Metrological Characterization of a Tunable Diode Laser Absorption Spectroscopy (TDLAS) Spectrometer for Simultaneous Absolute Measurement of Carbon Dioxide and Water Vapor. *Appl. Spectrosc.* **2015**, *69*, 257–268. [CrossRef]
10. Qu, Z.; Nwaboh, J.; Werhahn, O.; Ebert, V. Towards a dTDLAS-Based Spectrometer for Absolute HCl Measurements in Combustion Flue Gases and a Better Evaluation of Thermal Boundary Layer Effects. *Flow Turbul. Combust.* **2021**, *106*, 533–546. [CrossRef]
11. Wagner, S.; Klein, M.; Kathrotia, T.; Riedel, U.; Kissel, T.; Dreizler, A.; Ebert, V. Absolute, spatially resolved, in situ CO profiles in atmospheric laminar counter-flow diffusion flames using 2.3 μm TDLAS. *Appl. Phys. B* **2012**, *109*, 533–540. [CrossRef]
12. Wu, K.; Li, F.; Cheng, X.; Yang, Y.; Lin, X.; Xia, Y. Sensitive detection of CO<sub>2</sub> concentration and temperature for hot gases using quantum-cascade laser absorption spectroscopy near 4.2 μm. *Appl. Phys. B* **2014**, *117*, 659–666. [CrossRef]
13. Foltynowicz, A.; Schmidt, F.M.; Ma, W.; Axner, O. Noise-immune cavity-enhanced optical heterodyne molecular spectroscopy: Current status and future potential. *Appl. Phys. B* **2008**, *92*, 313–326. [CrossRef]
14. Henderson, B.; Khodabakhsh, A.; Metsälä, M.; Ventrillard, I.; Schmidt, F.M.; Romanini, D.; Ritchie, G.A.D.; Hekkert, S.T.L.; Briot, R.; Risby, T.; et al. Laser spectroscopy for breath analysis: Towards clinical implementation. *Appl. Phys. B* **2018**, *124*, 161. [CrossRef] [PubMed]
15. McCurdy, M.R.; Bakhirkin, Y.; Wysocki, G.; Lewicki, R.; Tittel, F.K. Recent advances of laser-spectroscopy-based techniques for applications in breath analysis. *J. Breath Res.* **2007**, *1*, 014001. [CrossRef] [PubMed]
16. Chao, X.; Jeffries, J.B.; Hanson, R.K. Absorption sensor for CO in combustion gases using 2.3 μm tunable diode lasers. *Meas. Sci. Technol.* **2009**, *20*, 115201. [CrossRef]
17. Goldenstein, C.S.; Spearrin, R.M.; Jeffries, J.B.; Hanson, R.K. Wavelength-modulation spectroscopy near 2.5 μm for H<sub>2</sub>O and temperature in high-pressure and -temperature gases. *Appl. Phys. B* **2014**, *116*, 705–716. [CrossRef]
18. Graf, M.; Scheidegger, P.; Kupferschmid, A.; Looser, H.; Peter, T.; Dirksen, R.; Emmenegger, L.; Tuzson, B. Compact and lightweight mid-infrared laser spectrometer for balloon-borne water vapor measurements in the UTLS. *Atmos. Meas. Tech.* **2021**, *14*, 1365–1378. [CrossRef]
19. Viciani, S.; Montori, A.; Chiarugi, A.; D’Amato, F. A Portable Quantum Cascade Laser Spectrometer for Atmospheric Measurements of Carbon Monoxide. *Sensors* **2018**, *18*, 2380. [CrossRef] [PubMed]
20. Sur, R.; Ding, Y.; Jackson, R.B.; Hanson, R.K. Tunable laser-based detection of benzene using spectrally narrow absorption features. *Appl. Phys. B* **2019**, *125*, 195. [CrossRef]
21. Liu, N.; Linguang, X.; Sheng, Z.; Tianbo, H.; Lei, Z.; Dianming, W.; Jingsong, L. Sensitive detection of atmospheric N<sub>2</sub>O isotopomers using a quantum cascade laser based spectrometer. *J. Quant. Spectrosc. Radiat. Transf.* **2019**, *236*, 106587. [CrossRef]
22. Sobanski, N.; Tuzson, B.; Scheidegger, P.; Looser, H.; Kupferschmid, A.; Iturrate, M.; Pascale, C.; Hüglin, C.; Emmenegger, L. Advances in High-Precision NO<sub>2</sub> Measurement by Quantum Cascade Laser Absorption Spectroscopy. *Appl. Sci.* **2021**, *11*, 1222. [CrossRef]
23. Diemel, O.; Honza, R.; Ding, C.-P.; Böhm, B.; Wagner, S. In situ sensor for cycle-resolved measurement of temperature and mole fractions in IC engine exhaust gases. *Proc. Combust. Inst.* **2019**, *37*, 1453–1460. [CrossRef]
24. Keabian, P.L.; Herndon, S.C.; Freedman, A. Detection of Nitrogen Dioxide by Cavity Attenuated Phase Shift Spectroscopy. *Anal. Chem.* **2005**, *77*, 724–728. [CrossRef] [PubMed]
25. Sur, R.; Peng, W.Y.; Strand, C.; Spearrin, R.M.; Jeffries, J.B.; Hanson, R.K.; Bekal, A.; Halder, P.; Poonacha, S.P.; Vartak, S.; et al. Mid-infrared laser absorption spectroscopy of NO<sub>2</sub> at elevated temperatures. *J. Quant. Spectrosc. Radiat. Transf.* **2017**, *187*, 364–374. [CrossRef]

26. Hundt, P.M.; Müller, M.; Mangold, M.; Tuzson, B.; Scheidegger, P.; Looser, H.; Hüglin, C.; Emmenegger, L. Mid-IR spectrometer for mobile, real-time urban NO<sub>2</sub> measurements. *Atmos. Meas. Tech.* **2018**, *11*, 2669–2681. [CrossRef]
27. Horii, C.V.; Zahniser, M.S.; Nelson, D.D., Jr.; McManus, J.B.; Wofsy, S.C. Nitric acid and nitrogen dioxide flux measurements: A new application of tunable diode laser absorption spectroscopy. In *Application of Tunable Diode and Other Infrared Sources for Atmospheric Studies and Industrial Processing Monitoring II*; SPIE: Denver, CO, USA, 1999; Volume 3758, pp. 152–161. [CrossRef]
28. Mazurenka, M.I.; Fawcett, B.L.; Elks, J.M.F.; Shallcross, D.E.; Orr-Ewing, A.J. 410-nm diode laser cavity ring-down spectroscopy for trace detection of NO<sub>2</sub>. *Chem. Phys. Lett.* **2003**, *367*, 1–9. [CrossRef]
29. Pan, Y.; Dong, L.; Yin, X.; Wu, H. Compact and Highly Sensitive NO<sub>2</sub> Photoacoustic Sensor for Environmental Monitoring. *Molecules* **2020**, *25*, 1201. [CrossRef] [PubMed]
30. Dooly, G.; Fitzpatrick, C.; Lewis, E. Hazardous exhaust gas monitoring using a deep UV based differential optical absorption spectroscopy (DOAS) system. *J. Phys. Conf. Ser.* **2007**, *76*, 012021. [CrossRef]
31. Jiang, Y.; Yang, T.; Liu, X.; Liu, Y. Automobile exhaust simulation detection system based on photoacoustic spectroscopy. In Proceedings of the ICOSM 2020: Optoelectronic Science and Materials, Hefei, China, 25–27 September 2020; Volume 11606, p. 1160608. [CrossRef]
32. Yamamoto, Y.; Sumizawa, H.; Yamada, H.; Tonokura, K. Real-time measurement of nitrogen dioxide in vehicle exhaust gas by mid-infrared cavity ring-down spectroscopy. *Appl. Phys. B* **2011**, *105*, 923–931. [CrossRef]
33. Li, Z.; Hu, R.; Xie, P.; Chen, H.; Liu, X.; Liang, S.; Wang, D.; Wang, F.; Wang, Y.; Lin, C.; et al. Simultaneous measurement of NO and NO<sub>2</sub> by a dual-channel cavity ring-down spectroscopy technique. *Atmos. Meas. Tech.* **2019**, *12*, 3223–3236. [CrossRef]
34. ISO. *ISO Guide 98-3, Guide to the Expression of Uncertainty in Measurement*; International Organization for Standardization: Geneva, Switzerland, 2008; ISBN 9267101889.
35. Zacharof, N.; Tietge, U.; Franco, V.; Mock, P. Type approval and real-world CO<sub>2</sub> and NO<sub>x</sub> emissions from EU light commercial vehicles. *Energy Policy* **2016**, *97*, 540–548. [CrossRef]
36. Werhahn, O.; Petersen, J.C. (Eds.) TILSAM Technical Protocol V1\_2010-09-29. 2010. Available online: [http://www.euramet.org/fileadmin/docs/projects/934\\_METCHEM\\_Interim\\_Report.pdf](http://www.euramet.org/fileadmin/docs/projects/934_METCHEM_Interim_Report.pdf) (accessed on 29 April 2021).
37. Nwaboh, J.A.; Persijn, S.; Arrhenius, K.; Bohlén, H.; Werhahn, O.; Ebert, V. Metrological quantification of CO in biogas using laser absorption spectroscopy and gas chromatography. *Meas. Sci. Technol.* **2018**, *29*, 095010. [CrossRef]
38. The NIST Standard Reference Photometer for Ozone Measurement Traceability. Available online: <https://www.nist.gov/programs-projects/nist-standard-reference-photometer-ozone-measurement-traceability> (accessed on 29 April 2021).
39. Witzel, O.; Klein, A.; Wagner, S.; Meffert, C.; Schulz, C.; Ebert, V. High-speed tunable diode laser absorption spectroscopy for sampling-free in-cylinder water vapor concentration measurements in an optical IC engine. *Appl. Phys. B* **2012**, *109*, 521–532. [CrossRef]
40. Klein, A.; Witzel, O.; Ebert, V. Rapid, Time-Division Multiplexed, Direct Absorption- and Wavelength Modulation-Spectroscopy. *Sensors* **2014**, *14*, 21497–21513. [CrossRef]
41. Witzel, O.; Klein, A.; Meffert, C.; Schulz, C.; Kaiser, S.A.; Ebert, V. Calibration-free, high-speed, in-cylinder laser absorption sensor for cycle-resolved, absolute H<sub>2</sub>O measurements in a production IC engine. *Proc. Combust. Inst.* **2015**, *35*, 3653–3661. [CrossRef]
42. Witzel, O.; Klein, A.; Meffert, C.; Wagner, S.; Kaiser, S.; Schulz, C.; Ebert, V. VCSEL-based, high-speed, in situ TDLAS for in-cylinder water vapor measurements in IC engines. *Opt. Express* **2013**, *21*, 19951. [CrossRef] [PubMed]
43. Nwaboh, J.A.; Qu, Z.; Li, G.; Kim, M.E.; Petersen, J.C.; Balslev-Harder, D.; Werhahn, O.; Ebert, V. Diode laser spectrometer for NO<sub>2</sub> quantification: Absolute laser spectroscopic and direct NO<sub>2</sub> concentration measurements for atmospheric monitoring within the EMPIR project MetNO<sub>2</sub>. In Proceedings of the EGU General Assembly 2020, Online. 4–8 May 2020. EGU2020-19259. [CrossRef]
44. Whiting, E.E. An empirical approximation to the Voigt profile. *J. Quant. Spectrosc. Radiat. Transf.* **1968**, *8*, 1379–1384. [CrossRef]
45. HITRAN Database. Available online: <https://hitran.org/> (accessed on 29 April 2021).
46. Ngo, N.H.; Lisak, D.; Tran, H.; Hartmann, J.-M. An isolated line-shape model to go beyond the Voigt profile in spectroscopic databases and radiative transfer codes. *J. Quant. Spectrosc. Radiat. Transf.* **2013**, *129*, 89–100. [CrossRef]
47. Ngo, N.H.; Lisak, D.; Tran, H.; Hartmann, J.-M. An isolated line-shape model to go beyond the Voigt profile in spectroscopic databases and radiative transfer codes. *J. Quant. Spectrosc. Radiat. Transf.* **2013**, *129*, 89–100, Erratum in *J. Quant. Spectrosc. Radiat. Transf.* **2014**, *134*, 105. [CrossRef]
48. Tennyson, J.; Bernath, P.F.; Campargue, A.; Csaszar, A.G.; Daumont, L.; Gamache, R.R.; Hodges, J.T.; Lisak, D.; Naumenko, O.V.; Rothman, L.S.; et al. Recommended isolated-line profile for representing high-resolution spectroscopic transitions (IUPAC Technical Report). *Pure Appl. Chem.* **2014**, *86*, 1931–1943. [CrossRef]
49. Werle, P.; Mücke, R.; Slemr, F. The limits of signal averaging in atmospheric trace-gas monitoring by tunable diode-laser absorption spectroscopy (TDLAS). *Appl. Phys. B* **1993**, *57*, 131–139. [CrossRef]
50. Deutscher Kalibrierdienst (DKD). Available online: <https://www.dkd.eu/> (accessed on 29 April 2021).
51. Werwein, V.; Brunzendorf, J.; Li, G.; Serdyukov, A.; Werhahn, O.; Ebert, V. High-resolution Fourier transform measurements of line strengths in the 00<sup>0</sup>2-00<sup>0</sup> main isotopologue band of nitrous oxide. *Appl. Opt.* **2017**, *56*, E99–E105. [CrossRef]
52. Pogany, A.; Ott, O.; Werhahn, O.; Ebert, V. Towards traceability in CO<sub>2</sub> line strength measurements by TDLAS at 2.7 μm. *J. Quant. Spectrosc. Radiat. Transf.* **2013**, *130*, 147–157. [CrossRef]