

Supplementary Data

Nanostructured Metal Oxide Sensors for Antibiotic Monitoring in Mineral and River Water

Cátia Magro ^{1,2,*}, Tiago Moura ², Joana Dionísio ¹, Paulo A. Ribeiro ³, Maria Raposo ³ and Susana Sério ^{3,*}

¹ School for International Training, World Learning Inc., Brattleboro, VT 05302, USA; joana.dionisio@sit.edu

² Department of Physics, NOVA School of Science and Technology, NOVA University Lisbon, 2829-516 Caparica, Portugal; ta.moura@campus.fct.unl.pt

³ Laboratory of Instrumentation, Biomedical Engineering and Radiation Physics (LIBPhys-UNL), Department of Physics, NOVA School of Science and Technology, NOVA University Lisbon, 2829-516 Caparica, Portugal; pfr@fct.unl.pt (P.A.R.); mfr@fct.unl.pt (M.R.)

* Correspondence: catia.magro@sit.edu (C.M.); susana.serio@fct.unl.pt (S.S.)

S1. Reproducibility Measurements of Figures S1–S6.

Table S1. Normalized Impedance data (Ω) reproducibility of the TiO_2 based sensors devices produced with 100% O_2 , when immersed in MW and RW at different AZI, CLA and ERY concentrations.

MW AZI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	0.0106	0.0039	0.0072	0.0034
10^{-14} M	0.0200	0.0073	0.0137	0.0063
10^{-13} M	0.0254	0.0094	0.0174	0.0080
10^{-12} M	0.0300	0.0117	0.0208	0.0091
10^{-11} M	0.0390	0.0150	0.0270	0.0120
10^{-10} M	0.0432	0.0173	0.0303	0.0129
10^{-9} M	0.0484	0.0187	0.0336	0.0148
10^{-8} M	0.0547	0.0207	0.0377	0.0170

MW CLARI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	0.0041	0.0056	0.0048	0.0008
10^{-14} M	0.0070	0.0106	0.0088	0.0018
10^{-13} M	0.0091	0.0133	0.0112	0.0021
10^{-12} M	0.0118	0.0155	0.0137	0.0018
10^{-11} M	0.0139	0.0234	0.0187	0.0048
10^{-10} M	0.0152	0.0285	0.0219	0.0067
10^{-9} M	0.0158	0.0213	0.0185	0.0028
10^{-8} M	0.0159	0.0230	0.0194	0.0035

MW ERIT				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	-0.007121814	0.003715304	-0.0017	0.0054
10^{-14} M	-0.010836905	0.008461669	-0.0012	0.0096
10^{-13} M	-0.014426082	0.011704099	-0.0014	0.0131
10^{-12} M	-0.017404262	0.013934241	-0.0017	0.0157
10^{-11} M	-0.017649471	0.018165389	0.0003	0.0179
10^{-10} M	-0.017293276	0.021319272	0.0020	0.0193
10^{-9} M	-0.02298693	0.023543261	0.0003	0.0233
10^{-8} M	-0.020006048	0.02746048	0.0037	0.0237

10^{-7} M	0.0591	0.0218	0.0404	0.0187
10^{-6} M	0.0643	0.0244	0.0443	0.0200
10^{-5} M	0.0706	0.0271	0.0489	0.0217

10^{-7} M	0.0181	0.0546	0.0364	0.0182
10^{-6} M	0.0194	0.0746	0.0470	0.0276
10^{-5} M	0.0192	0.0321	0.0256	0.0065

10^{-7} M	-0.021724946	0.030946648	0.0046	0.0263
10^{-6} M	-0.021508108	0.03403566	0.0063	0.0278
10^{-5} M	-0.01754882	0.039208488	0.0108	0.0284

RW AZI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	-0.0126	-0.0007	-0.0066	0.0060
10^{-14} M	-0.0306	-0.0023	-0.0165	0.0141
10^{-13} M	-0.0529	-0.0029	-0.0279	0.0250
10^{-12} M	-0.0773	-0.0040	-0.0406	0.0367
10^{-11} M	-0.0971	-0.0048	-0.0510	0.0462
10^{-10} M	-0.1205	-0.0066	-0.0635	0.0570
10^{-9} M	-0.1480	-0.0092	-0.0786	0.0694
10^{-8} M	-0.1767	-0.0098	-0.0932	0.0835
10^{-7} M	-0.2013	-0.0116	-0.1065	0.0948
10^{-6} M	-0.2252	-0.0150	-0.1201	0.1051
10^{-5} M	-0.2454	-0.0155	-0.1305	0.1150

RW CLARI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	-0.0383	-0.0160	-0.0272	0.0112
10^{-14} M	-0.0576	-0.0276	-0.0426	0.0150
10^{-13} M	-0.0732	-0.0371	-0.0551	0.0180
10^{-12} M	-0.0866	-0.0455	-0.0661	0.0206
10^{-11} M	-0.1125	-0.0588	-0.0857	0.0269
10^{-10} M	-0.1251	-0.0673	-0.0962	0.0289
10^{-9} M	-0.1415	-0.0764	-0.1090	0.0326
10^{-8} M	-0.1581	-0.0862	-0.1221	0.0360
10^{-7} M	-0.1784	-0.0956	-0.1370	0.0414
10^{-6} M	-0.1943	-0.1100	-0.1521	0.0422
10^{-5} M	-0.2144	-0.1140	-0.1642	0.0502

RW ERIT				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	0.0041	0.0120	0.0080	0.0039
10^{-14} M	0.0043	0.0204	0.0123	0.0081
10^{-13} M	0.0094	0.0277	0.0186	0.0091
10^{-12} M	0.0107	0.0356	0.0231	0.0124
10^{-11} M	0.0142	0.0477	0.0310	0.0168
10^{-10} M	0.0180	0.0547	0.0363	0.0183
10^{-9} M	0.0212	0.0609	0.0411	0.0198
10^{-8} M	0.0234	0.0698	0.0466	0.0232
10^{-7} M	0.0274	0.0767	0.0520	0.0246
10^{-6} M	0.0308	0.0846	0.0577	0.0269
10^{-5} M	0.0325	0.0887	0.0606	0.0281

Table S2. Normalized Impedance data (Ω) reproducibility of the TiO_2 based sensors devices produced with 50% O_2 , when immersed in MW and RW at different AZI, CLA and ERY concentrations.

MW AZI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	-0.5355	-0.2953	-0.4154	0.1201
10^{-14} M	-0.5394	-0.3480	-0.4437	0.0957
10^{-13} M	-0.5986	-0.4575	-0.5280	0.0705
10^{-12} M	-0.6275	-0.4443	-0.5359	0.0916

MW CLARI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	-0.3162	-0.2142	-0.2652	0.0510
10^{-14} M	-0.3885	-0.3214	-0.3550	0.0336
10^{-13} M	-0.4492	-0.4107	-0.4299	0.0193
10^{-12} M	-0.4677	-0.4700	-0.4688	0.0012

MW ERIT				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	-0.298638445	-0.289563448	-0.2941	0.0045
10^{-14} M	-0.375538138	-0.388109988	-0.3818	0.0063
10^{-13} M	-0.481711642	-0.452344157	-0.4670	0.0147
10^{-12} M	-0.533672326	-0.49143296	-0.5126	0.0211

10^{-11} M	-0.6473	-0.4753	-0.5613	0.0860
10^{-10} M	-0.6793	-0.5020	-0.5907	0.0887
10^{-9} M	-0.7241	-0.5479	-0.6360	0.0881
10^{-8} M	-0.7270	-0.5337	-0.6304	0.0967
10^{-7} M	-0.7520	-0.5442	-0.6481	0.1039
10^{-6} M	-0.7631	-0.5455	-0.6543	0.1088
10^{-5} M	-0.7644	-0.5480	-0.6562	0.1082

10^{-11} M	-0.4981	-0.5585	-0.5283	0.0302
10^{-10} M	-0.5452	-0.6201	-0.5827	0.0375
10^{-9} M	-0.5430	-0.6624	-0.6027	0.0597
10^{-8} M	-0.5660	-0.6192	-0.5926	0.0266
10^{-7} M	-0.6192	-0.6621	-0.6406	0.0215
10^{-6} M	-0.5851	-0.6835	-0.6343	0.0492
10^{-5} M	-0.6135	-0.7013	-0.6574	0.0439

10^{-11} M	-0.543996559	-0.514074815	-0.5290	0.0150
10^{-10} M	-0.580831031	-0.526168915	-0.5535	0.0273
10^{-9} M	-0.612072783	-0.55132712	-0.5817	0.0304
10^{-8} M	-0.620737614	-0.566044259	-0.5934	0.0273
10^{-7} M	-0.62728041	-0.567301508	-0.5973	0.0300
10^{-6} M	-0.653309718	-0.604771693	-0.6290	0.0243
10^{-5} M	-0.669442648	-0.586878472	-0.6282	0.0413

RW AZI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	-0.5100	-0.4319	-0.4709	0.0390
10^{-14} M	-0.6604	-0.4829	-0.5716	0.0887
10^{-13} M	-0.7097	-0.5731	-0.6414	0.0683
10^{-12} M	-0.7579	-0.6480	-0.7030	0.0550
10^{-11} M	-0.7888	-0.6509	-0.7199	0.0690
10^{-10} M	-0.7884	-0.6783	-0.7334	0.0550
10^{-9} M	-0.8004	-0.6726	-0.7365	0.0639
10^{-8} M	-0.8049	-0.6930	-0.7490	0.0560
10^{-7} M	-0.8108	-0.6932	-0.7520	0.0588
10^{-6} M	-0.8103	-0.7073	-0.7588	0.0515
10^{-5} M	-0.8150	-0.7093	-0.7622	0.0529

RW CLARI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	-0.0441	-0.0396	-0.0419	0.0023
10^{-14} M	-0.0575	-0.0480	-0.0528	0.0048
10^{-13} M	-0.0603	-0.0531	-0.0567	0.0036
10^{-12} M	-0.0589	-0.0568	-0.0578	0.0010
10^{-11} M	-0.0590	-0.0562	-0.0576	0.0014
10^{-10} M	-0.0598	-0.0566	-0.0582	0.0016
10^{-9} M	-0.0632	-0.0595	-0.0614	0.0018
10^{-8} M	-0.0642	-0.0596	-0.0619	0.0023
10^{-7} M	-0.0658	-0.0603	-0.0630	0.0027
10^{-6} M	-0.0675	-0.0613	-0.0644	0.0031
10^{-5} M	-0.0689	-0.0603	-0.0646	0.0043

RW ERIT				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	-0.4258	-0.4472	-0.4365	0.0107
10^{-14} M	-0.6010	-0.5913	-0.5962	0.0048
10^{-13} M	-0.6495	-0.7111	-0.6803	0.0308
10^{-12} M	-0.6773	-0.7205	-0.6989	0.0216
10^{-11} M	-0.6977	-0.7394	-0.7186	0.0208
10^{-10} M	-0.7129	-0.7539	-0.7334	0.0205
10^{-9} M	-0.7255	-0.7654	-0.7455	0.0200
10^{-8} M	-0.7361	-0.7697	-0.7529	0.0168
10^{-7} M	-0.7415	-0.7731	-0.7573	0.0158
10^{-6} M	-0.7577	-0.7777	-0.7677	0.0100
10^{-5} M	-0.7615	-0.7819	-0.7717	0.0102

Table S3. Normalized Impedance data (Ω) reproducibility of the ZnObased sensors devices produced with 100% O₂, when immersed in MW and RW at different AZI, CLA and ERY concentrations.

MW AZI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10 ⁻¹⁵ M	-0.0932	0.3406	0.1237	0.2169
10 ⁻¹⁴ M	-0.1248	0.4380	0.1566	0.2814
10 ⁻¹³ M	-0.2085	0.4313	0.1114	0.3199
10 ⁻¹² M	-0.2404	0.3799	0.0697	0.3101
10 ⁻¹¹ M	-0.2473	0.3872	0.0699	0.3172
10 ⁻¹⁰ M	-0.2714	0.3721	0.0503	0.3217
10 ⁻⁹ M	-0.3060	0.3234	0.0087	0.3147
10 ⁻⁸ M	-0.3372	0.2931	-0.0220	0.3151
10 ⁻⁷ M	-0.3450	0.2608	-0.0421	0.3029
10 ⁻⁶ M	-0.3723	0.2018	-0.0853	0.2870
10 ⁻⁵ M	-0.3768	0.2149	-0.0809	0.2959

MW CLARI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10 ⁻¹⁵ M	0.2356	1.2696	0.7526	0.5170
10 ⁻¹⁴ M	0.5043	3.0169	1.7606	1.2563
10 ⁻¹³ M	0.7534	5.6357	3.1946	2.4411
10 ⁻¹² M	0.9321	8.8369	4.8845	3.9524
10 ⁻¹¹ M	1.1118	12.6809	6.8964	5.7846
10 ⁻¹⁰ M	1.3347	17.8506	9.5926	8.2579
10 ⁻⁹ M	1.5415	23.7623	12.6519	11.1104
10 ⁻⁸ M	1.7610	30.4630	16.1120	14.3510
10 ⁻⁷ M	1.9612	36.7914	19.3763	17.4151
10 ⁻⁶ M	2.1190	47.1884	24.6537	22.5347
10 ⁻⁵ M	2.3429	53.0738	27.7083	25.3654

MW ERIT				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10 ⁻¹⁵ M	-0.015635615	0.038991603	0.0117	0.0273
10 ⁻¹⁴ M	-0.100054729	0.022152197	-0.0390	0.0611
10 ⁻¹³ M	-0.175895981	-0.049136708	-0.1125	0.0634
10 ⁻¹² M	-0.21358818	-0.035794482	-0.1247	0.0889
10 ⁻¹¹ M	-0.236529308	-0.021220127	-0.1289	0.1077
10 ⁻¹⁰ M	-0.271723587	-0.064155612	-0.1679	0.1038
10 ⁻⁹ M	-0.281178095	-0.076722196	-0.1790	0.1022
10 ⁻⁸ M	-0.323372253	-0.097834268	-0.2106	0.1128
10 ⁻⁷ M	-0.341931737	-0.157576559	-0.2498	0.0922
10 ⁻⁶ M	-0.365188325	-0.144436119	-0.2548	0.1104
10 ⁻⁵ M	-0.391553519	-0.178290771	-0.2849	0.1066

RW AZI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10 ⁻¹⁵ M	-0.2076	-0.1596	-0.1836	0.0240
10 ⁻¹⁴ M	-0.2727	-0.2153	-0.2440	0.0287
10 ⁻¹³ M	-0.3113	-0.2844	-0.2979	0.0134
10 ⁻¹² M	-0.3346	-0.3097	-0.3222	0.0124
10 ⁻¹¹ M	-0.3527	-0.3283	-0.3405	0.0122
10 ⁻¹⁰ M	-0.3658	-0.3463	-0.3561	0.0097
10 ⁻⁹ M	-0.3869	-0.3613	-0.3741	0.0128

RW CLARI				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10 ⁻¹⁵ M	0.0270	-0.1053	-0.0392	0.0662
10 ⁻¹⁴ M	-0.0222	-0.1742	-0.0982	0.0760
10 ⁻¹³ M	-0.0678	-0.2171	-0.1425	0.0746
10 ⁻¹² M	-0.1209	-0.2542	-0.1876	0.0666
10 ⁻¹¹ M	-0.1415	-0.2771	-0.2093	0.0678
10 ⁻¹⁰ M	-0.2003	-0.2979	-0.2491	0.0488
10 ⁻⁹ M	-0.2159	-0.3188	-0.2674	0.0514

RW ERIT				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation
10 ⁻¹⁵ M	-0.0106	-0.0199	-0.0152	0.0046
10 ⁻¹⁴ M	-0.0168	-0.0306	-0.0237	0.0069
10 ⁻¹³ M	-0.0196	-0.0365	-0.0280	0.0084
10 ⁻¹² M	-0.0268	-0.0429	-0.0348	0.0080
10 ⁻¹¹ M	-0.0310	-0.0629	-0.0469	0.0160
10 ⁻¹⁰ M	-0.0347	-0.0593	-0.0470	0.0123
10 ⁻⁹ M	-0.0381	-0.0601	-0.0491	0.0110

10^{-8} M	-0.3907	-0.3726	-0.3817	0.0090		10^{-8} M	-0.2211	-0.3363	-0.2787	0.0576		10^{-8} M	-0.0397	-0.0644	-0.0521	0.0123
10^{-7} M	-0.4047	-0.3869	-0.3958	0.0089		10^{-7} M	-0.2422	-0.3592	-0.3007	0.0585		10^{-7} M	-0.0429	-0.0735	-0.0582	0.0153
10^{-6} M	-0.4097	-0.4065	-0.4081	0.0016		10^{-6} M	-0.2646	-0.3770	-0.3208	0.0562		10^{-6} M	-0.0416	-0.0725	-0.0571	0.0155
10^{-5} M	-0.4167	-0.4264	-0.4215	0.0049		10^{-5} M	-0.2740	-0.3916	-0.3328	0.0588		10^{-5} M	-0.0446	-0.0778	-0.0612	0.0166

Table S4. Normalized Impedance data (Ω) reproducibility of the ZnO based sensors devices produced with 50% O₂, when immersed in MW and RW at different AZI, CLA and ERY concentrations.

MW AZI					MW CLARI					MW ERIT				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation	Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation	Concentration s	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	0.0117	-0.0077	0.0020	0.0097	10^{-15} M	0.0108	0.0040	0.0074	0.0034	10^{-15} M	0.0418	0.0171	0.0295	0.0123
10^{-14} M	0.0276	-0.0084	0.0096	0.0180	10^{-14} M	0.0277	0.0126	0.0201	0.0075	10^{-14} M	0.1028	0.0544	0.0786	0.0242
10^{-13} M	0.0443	-0.0083	0.0180	0.0263	10^{-13} M	0.0412	0.0224	0.0318	0.0094	10^{-13} M	0.1482	0.0806	0.1144	0.0338
10^{-12} M	0.0555	-0.0086	0.0235	0.0320	10^{-12} M	0.0531	0.0304	0.0418	0.0113	10^{-12} M	0.1872	0.1059	0.1465	0.0407
10^{-11} M	0.0671	-0.0071	0.0300	0.0371	10^{-11} M	0.0629	0.0392	0.0510	0.0118	10^{-11} M	0.2282	0.1266	0.1774	0.0508
10^{10} M	0.0789	-0.0055	0.0367	0.0422	10^{-10} M	0.0750	0.0480	0.0615	0.0135	10^{-10} M	0.2638	0.1464	0.2051	0.0587
10^{-9} M	0.0873	-0.0037	0.0418	0.0455	10^{-9} M	0.0864	0.0558	0.0711	0.0153	10^{-9} M	0.2978	0.1631	0.2304	0.0674
10^{-8} M	0.0985	-0.0005	0.0490	0.0495	10^{-8} M	0.0965	0.0633	0.0799	0.0166	10^{-8} M	0.3347	0.1857	0.2602	0.0745
10^{-7} M	0.1100	-0.0001	0.0550	0.0551	10^{-7} M	0.1068	0.0765	0.0917	0.0151	10^{-7} M	0.3654	0.1979	0.2816	0.0837
10^{-6} M	0.1195	0.0023	0.0609	0.0586	10^{-6} M	0.1179	0.0839	0.1009	0.0170	10^{-6} M	0.4034	0.2153	0.3094	0.0940
10^{-5} M	0.1287	0.0066	0.0677	0.0611	10^{-5} M	0.1278	0.0911	0.1094	0.0183	10^{-5} M	0.4290	0.2257	0.3273	0.1017

RW AZI					RW CLARI					RW ERIT				
Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation	Concentrations	Sensor 1	Sensor 2	Average	Standard Deviation	Concentration s	Sensor 1	Sensor 2	Average	Standard Deviation
10^{-15} M	0.0396	0.0213	0.0305	0.0091	10^{-15} M	0.0067	0.0014	0.0040	0.0026	10^{-15} M	0.0365	0.0107	0.0236	0.0129
10^{-14} M	0.0248	0.0282	0.0265	0.0017	10^{-14} M	0.0380	0.0035	0.0207	0.0172	10^{-14} M	0.0520	0.0191	0.0355	0.0165
10^{-13} M	0.0114	0.0275	0.0194	0.0081	10^{-13} M	0.0701	0.0059	0.0380	0.0321	10^{-13} M	0.0598	0.0215	0.0407	0.0192

MW AZI				
10^{-12} M	0.0123	0.0274	0.0199	0.0075
10^{-11} M	0.0154	0.0296	0.0225	0.0071
10^{-10} M	0.0163	0.0310	0.0237	0.0073
10^{-9} M	0.0176	0.0330	0.0253	0.0077
10^{-8} M	0.0196	0.0343	0.0269	0.0073
10^{-7} M	0.0253	0.0358	0.0306	0.0053
10^{-6} M	0.0241	0.0371	0.0306	0.0065
10^{-5} M	0.0256	0.0390	0.0323	0.0067

MW CLARI				
10^{-12} M	0.1441	0.0104	0.0773	0.0668
10^{-11} M	0.1876	0.0131	0.1003	0.0873
10^{-10} M	0.2149	0.0176	0.1163	0.0987
10^{-9} M	0.2387	0.0215	0.1301	0.1086
10^{-8} M	0.2630	0.0302	0.1466	0.1164
10^{-7} M	0.2751	0.0341	0.1546	0.1205
10^{-6} M	0.2878	0.0375	0.1626	0.1252
10^{-5} M	0.3007	0.0404	0.1706	0.1302

MW ERIT				
10^{-12} M	0.0642	0.0243	0.0443	0.0200
10^{-11} M	0.0685	0.0254	0.0470	0.0216
10^{-10} M	0.0729	0.0283	0.0506	0.0223
10^{-9} M	0.0775	0.0273	0.0524	0.0251
10^{-8} M	0.0810	0.0279	0.0544	0.0265
10^{-7} M	0.0847	0.0329	0.0588	0.0259
10^{-6} M	0.0892	0.0422	0.0657	0.0235
10^{-5} M	0.0907	0.0377	0.0642	0.0265




Figure S1. Normalized impedance as a function of azithromycin concentration in MW at fixed frequencies: 1, 1.58, 2.51 and 2.51×10^3 Hz measured by the TiO_2 (100% O_2) sensor, TiO_2 (50% O_2) sensor, ZnO (100% O_2) sensor, and ZnO (50% O_2) sensor, respectively. To note that the percentage inside the parentheses correspond to the oxygen percentage used to reactive sputtered the corresponding metallic target.




Figure S2. Normalized impedance as a function of azithromycin concentration in RW at fixed frequencies: 1, 1.58 , 3.98×10^3 and 6.31×10^4 Hz measured by TiO₂ (50% O₂) sensor, ZnO (100% O₂) sensor, TiO₂ (100% O₂) sensor, and ZnO (50% O₂) sensor, respectively. To note that the percentage inside the parentheses correspond to the oxygen percentage used to reactive sputtered the corresponding metallic target.




Figure S3. Normalized impedance as a function of clarithromycin concentration in MW at fixed frequencies: 1, 1, 2.51, 2.51 Hz measured by the TiO₂ (50% O₂) sensor, TiO₂ (100% O₂) sensor, ZnO (50% O₂) sensor and ZnO (100% O₂) sensor, respectively. To note that the percentage inside the parentheses correspond to the oxygen percentage used to reactive sputtered the corresponding metallic target.




Figure S4. Normalized impedance as a function of clarithromycin concentration in RW at fixed frequencies: 1, 3.98×10^2 , 2.51×10^3 and 10^6 Hz measured by the ZnO (100% O₂) sensor, ZnO (50% O₂) sensor, TiO₂ (100% O₂) sensor, and TiO₂ (50% O₂) sensor, respectively. To note that the percentage inside the parentheses correspond to the oxygen percentage used to reactive sputtered the corresponding metallic target.




Figure S5. Normalized impedance as a function of erythromycin concentration in MW at fixed frequencies: 1, 1.58×10^2 and 2.51×10^3 Hz measured by the TiO₂ (50% O₂) sensor, ZnO (50% O₂) sensor, ZnO (100% O₂), TiO₂ (100% O₂) sensor, respectively. To note that the percentage inside the parentheses correspond to the oxygen percentage used to reactive sputtered the corresponding metallic target.






Figure S6. Normalized impedance as a function of erythromycin concentration in RW at fixed frequencies: 1, 10, 2.51×10^4 and 10^6 Hz measured by the TiO₂ (50% O₂), TiO₂ (100% O₂) sensor, ZnO (50% O₂) sensor, and ZnO (100% O₂) sensor, respectively. To note that the percentage inside the parentheses correspond to the oxygen percentage used to reactive sputtered the corresponding metallic target.



(a) ZnO film deposited with 50% O₂



(b) ZnO film deposited with 100% O₂




Figure S7. SEM images with 30,000 times magnification for (a) ZnO 50%; (b) ZnO 100%; (c) TiO₂ 50%; (d) TiO₂ 100% (scale bar size: 5 μ m).