Development of Electrochemical Sensors Based on Electrosynthesized Ion Imprinted Polymers for Cobalt (Co^{2+}) Ions Determination in Water

Nelson Manrique-Rodríguez, Sabrina Di Masi* and Cosimino Malitesta

Laboratorio di Chimica Analitica, Dipartimento di Scienze e Tecnologie Biologiche ed Ambientali, Lecce, Italia.

UNIVERSITÀ DEL SALENTO
L'Ateneo tra i due mari

*sabrina.dimasi@unisalento.it

CONTENT

- ✓ INTRODUCTION
- ✓ OBJECTIVE
- ✓ METHODS
- ✓ RESULTS
- ✓ CONCLUSIONS

INTRODUCTION



- ✓ COBALT POLLUTION IN SOIL AND WATER
- ✓ MINING INDUSTRY, DOMESTIC AND INDUSTRIAL WASTEWATER, BURNING OF FOSSIL FUELS
- ✓ AFFECTS GROWTH AND YIELD OF AGRICULTURE CROP PLANTS
- ✓ POTENTIAL RISK TO FOOD SOURCES AND HUMAN HEALTH
- ✓ MONITOR EVEN TRACE AMOUNTS OF COBALT



- ✓ ATOMIC ABSORTION SPECTROMETRY
- ✓ INDUCTIVELY COUPLED PLASMA MASS SPECTROMETRY
- ✓ FLUORESCENCE METHODS
- ✓ ELECTROCHEMICAL TECHNIQUES
- ✓ INTERFERENCE, SAMPLE PREPARATION, BEING TIME-CONSUMING, EXPENSIVE INSTRUMENTS, AND LOW SENSITIVITY.
- ✓ SIMPLE, RAPID, SENSITIVE, AND SPECIFIC METHOD FOR DETECTING Co^{2+}







CHEMICAL SENSOR: A SCHEME



OUTPUT (SENSOR RESPONSE)



RESPONSE EXAMPLE OF A VOLTAMMETRIC SENSOR

AN HIGH SELECTIVE RECOGNITION ELEMENTS: ION IMPRINTED POLYMER (IIP)

CHEMICAL SYNTHESIS



ELECTROCHEMICAL SYNTHESIS



DEVELOPMENT OF AN ELECTROCHEMICAL IMPEDIMETRIC SENSOR BASED ON AN ION IMPRINTED POLYMERIC FILM FOR THE DETERMINATION OF Co^{2+} IONS IN WATER OF ENVIRONMENTAL INTEREST



2-aminophenol monomer

Co(II) ions

Drop-casting



Screen-printed carbon electrodes, SPCE

Cyclic voltammetry

Washing step



Electrochemical synthesis of ion imprinted polymeric film based on polyaminophenol (Co(II)-IIP film) and not imprinted polymeric film (NIP film, control)

Electrochemical Impedance Spectroscopy



Analytical performances of IIP film and NIP film

METHODS

Ion Imprinted Polymer (IIP) Preparation and Characterization



PalmSens Compact Electrochemical Interfaces
PSTrace 5

CYCLIC VOLTAMMETRY CV
ELECTROCHEMICAL IMPEDANCE ESPECTROSCOPY EIS

SCREEN PRINTED CARBON ELECTRODE | IIP | 2-AP | Co^{2+} | PBS

CV POLIMERIZATION



OPTIMISATION BY MULTIVARIATE EXPERIMENTAL DESIGN

MODDE®

SARTORIUS

DoE

23 different experiments, 4 affecting variables

Y = Sensitivity of IIP sensor

Experiment number	Experiment name	Run Order	Incl/Excl	2-AP concentration	CO^{2+} concentration	Cycles	Time elution
1	N1	18	Incl	1	1	10	10
2	N2	1	Incl	2	1	10	10
3	N3	5	Incl	1	5	10	10
4	N4	2	Incl	2	1	30	10
5	N5	23	Incl	1	5	30	10
6	N6	22	Incl	2	5	30	10
7	N7	3	Incl	1	1	10	25
8	N8	9	Incl	1	5	10	25
9	N9	6	Incl	2	5	10	25
10	N10	8	Incl	1	1	30	25
11	N11	16	Incl	2	1	30	25
12	N12	11	Incl	2	5	30	25
13	N13	19	Incl	1	3	20	17.5
14	N14	15	Incl	2	3	20	17.5
15	N15	20	Incl	1.5	1	20	17.5
16	N16	12	Incl	1.5	5	20	17.5
17	N17	4	Incl	1.5	3	10	17.5
18	N18	17	Incl	1.5	3	30	17.5
19	N19	7	Incl	1.5	3	20	10
20	N20	14	Incl	1.5	3	20	25
21	N21	13	Incl	1.5	3	20	17.5
22	N22	10	Incl	1.5	3	20	17.5
23	N23	21	Incl	1.5	3	20	17.5

CONCLUSIONS

- ✓ Adequate properties of selectivity and stability of the IIP films, compared with the yields obtained with other sensors.
- ✓ 44% of experiments with best fit comparing circuit model vs. experimental data.
- ✓ Sensitivity of Co(II)-IIP film sensor towards Co^{2+} ions in the range of concentration between 1.9 and 32.5 µmol L^{-1} was 5 times higher than that reported for NIP film, revealing the goodness of imprinted process.

UNIVERSITÀ DEL SALENTO

L'Ateneo tra i due mari

Thank you