

Abstract

Using Interdigitated Organic Electrochemical Transistors as Electrophysiological and Biochemical Sensors [†]

Dirk Mayer 

Institute of Biological Information Processing, Bioelectronics (IBI-3), Forschungszentrum Jülich, 52425 Jülich, Germany; dirk.mayer@fz-juelich.de

[†] Presented at the 8th International Symposium on Sensor Science, 17–28 May 2021; Available Online: <https://i3s2021dresden.sciforum.net/>.

Abstract: Organic electrochemical transistors (OECTs) have emerged as versatile electrophysiological sensors due to their high transconductance, biocompatibility, and transparent channel material. High maximum transconductances were demonstrated, facilitating extracellular recordings from electrogenic cells. However, this often requires large channel dimensions which impede high transistor densities. To improve the device performance and density, we used interdigitated OECTs (iOECTs), which feature high transconductances with small device areas. Superior device performance was achieved by systematically optimizing the electrode layout regarding channel length, number of electrode digits, and electrode width. Interestingly, the maximum transconductance does not straightforwardly scale with the channel width-to-length ratio, which is different from planar OECTs. We used optimized iOECTs for recording action potentials of cardiomyocyte-like HL-1 cells. Furthermore, we embedded the iOECTs in a matrix of polyimide to achieve flexible and transparent bioelectronic devices. These sensors exhibited electrical characteristics similar to those of solid-substrate devices even after experiencing extremely high bending strain. Finally, we used these devices to detect neurotransmitter dopamine and ATP, which play an important role not only in signal transmission in the central nervous system but also in cardiovascular, neurodegenerative, and immune deficiency diseases. Our novel aptasensor possessed ultralow detection limits, which were several orders of magnitude lower than those of the same aptasensors using an amperometric transducer principle. Our results demonstrate that interdigitated OECTs meet two requirements of both electrophysiological and biochemical sensors, namely high device performance and small channel dimensions, and might represent the optimal transducer to integrate these two types of sensors on one chip.



Citation: Mayer, D. Using Interdigitated Organic Electrochemical Transistors as Electrophysiological and Biochemical Sensors. *Eng. Proc.* **2021**, *6*, 46. <https://doi.org/10.3390/I3S2021Dresden-10133>

Academic Editors: Gianarelio Cuniberti and Larysa Baraban

Published: 17 May 2021

Publisher's Note: MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Keywords: interdigitated organic electrochemical transistors; aptasensors; dopamine; ATP; electrophysiology

Supplementary Materials: The supplementary file is available online at <https://www.mdpi.com/article/10.3390/I3S2021Dresden-10133/s1>, Poster S1: Dirk Mayer I3S2021 OECT kurz-2_21cecc4e-5bb9-4c80-aec1-edd0c75f271b.



Copyright: © 2021 by the author. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).